

CERTIFICATE OF COMPLIANCE

Certificate Number 2022-08-15-E484517
Report Reference E484517-D1001-1/A0/C0-UL
Date 2022-08-15

Issued to: DFI Inc
Applicant Company: 10F No 97 Sec 1 Xintai 5th Rd Xizhi District
New Taipei, 22175 Taiwan TAIWAN

Listed Company: Same as applicant

This is to certify that representative samples of Healthcare Display
MDP156

Have been investigated by UL in accordance with the Standard(s) indicated on this Certificate.

Standard(s) for Safety: ANSI/AAMI ES60601-1:2005/(R)2012 and A1:2012,
C1:2009/(R)2012 and A2:2010/(R)2012, CSA CAN/CSA-C22.2
NO. 60601-1:14

Additional Standards: N/A

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information.

This Certificate of Compliance does not provide authorization to apply the UL Mark. Only the UL Follow-Up Services Procedure provides authorization to apply the UL Mark.

Only those products bearing the UL Mark should be considered as being UL Certified and covered under UL's Follow-Up Services.

Look for the UL Certification Mark on the product.



Bruce Mahrenholz, Director North American Certification Program
UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL Customer Service Representative at <http://ul.com/aboutul/locations/>

